



State Hygienic Laboratory

The University of Iowa

JON MARTENS
ATLANTIC MUNICIPAL UTILITIES
15 W 3RD ST
ATLANTIC, IA 50022-

Accession Number	23981
Date Sample Finalized	2012-04-06 13:13
Date Received	2012-04-05 09:30
Sample Source	Drinking Water
Project	
Date Collected	2012-04-04 10:10
Collection Site	well #6
Collection Town	ATLANTIC
Sample Description	water
Client Reference	
Collector	parrott troy
Phone	712/243-1395

Note: Upon arrival, sample met container and preservation requirements for the analysis requested. Please review carefully your sample results for additional analyte comments or method exceptions.

Results of Analyses

GCMS Volatiles, EPA 524.2

Units	mg/L
Date Analyzed	2012-04-05 12:30
Analyst	LL

Analyzed In	Iowa City
Date Verified	2012-04-06 13:13
Verifier	WB

Analyte	Result	Quant Limit	MCL
Tetrachloroethylene	0.0067	0.0005	0.005

Note: The MCL (maximum contaminant level) is only applicable to compliance monitoring samples under the Safe Drinking Water Act (SDWA).

Description of Units used within this report

mg/L = Milligrams per Liter

The result(s) of this report relate only to the items analyzed. This report shall not be reproduced except in full without the written approval of the laboratory.

Iowa Environmental Laboratory IDs are: Ankeny #397, Iowa City/Coralville #027, Lakeside #393.

If you have any questions, please call Client Services at 800/421-IOWA (4692) or 319/335-4500. Thank you.

A72Q

40479921

1-0



Superfund

OU 00

Page 1 of 1

Michael D. Wichman, Ph.D.
Michael A. Pentella, Ph.D.
Associate Directors
<http://www.shl.uiowa.edu>

University of Iowa Research Park
2490 Crosspark Road
Coralville, IA 52241
319/335-4500 Fax: 319/335-4555

Lakeside Laboratory
1838 Highway 86
Milford, IA 51351
712/337-3669 ext. 6 Fax: 712/337-0227

Iowa Laboratories Complex
2220 S. Ankeny Blvd
Ankeny, IA 50023
515/725-1600 Fax: 515/725-1642